

01 July 2024

To: Hospitals and surgeons

**Subject: URGENT MEDICAL DEVICE FIELD SAFETY NOTICE (CORRECTION)
INCREASED RISK OF PERIPROSTHETIC FEMORAL FRACTURE**

Affected Product: CPT® Hip System Femoral Stem 12/14 Neck Taper (see **Attachment 2** for the **Affected Product List**)



Reason for Field Safety Corrective Action

Zimmer, Inc. is voluntarily initiating a medical device Field Safety Corrective Action (correction) for the CPT Hip System Femoral Stem 12/14 Neck Taper products listed in **Attachment 2 – Affected Product List** due to an increased risk of postoperative periprosthetic femoral fracture (PFF).

You are receiving this Field Safety Notice because our records indicate that you have implanted the CPT Hip System Femoral Stem 12/14 Neck Taper in the past 12 months, or your facility may currently have inventory of this product.

The CPT Hip System Femoral Stem 12/14 Neck Taper is a polished taper-slip (PTS) style stem. Though PTS stems demonstrate low rates of aseptic loosening, contributing to high rates of survivorship, registry-based analyses have indicated that all PTS stems are associated with increased risk of PFF compared to composite beam stems.^{3,4} Within the PTS stem category, stems manufactured from cobalt chromium alloy are at increased risk for PFF in comparison to stainless steel.⁵ The CPT Hip System Femoral Stem 12/14 Neck Taper is manufactured from cobalt chromium alloy.

While PFF is a generally recognized risk associated with all total hip arthroplasty, the current instructions for use (IFU) for the CPT Hip System Femoral Stem 12/14 Neck Taper do not identify post-operative bone fracture as a risk. As a result, Zimmer, Inc. is initiating a Field Safety Corrective Action to ensure that orthopaedic surgeons are aware of this risk and that the CPT Hip System Femoral Stem 12/14 Neck Taper displayed a rate of PFF at approximately 1.4% in the United Kingdom, which is more than twice the risk of PFF relative to a similar PTS reference product that is stainless steel.⁶ The IFU is in the process of being updated to reflect the risk of PFF, with an anticipated release by the end of August 2024.

Risks

In the event that PFF occurs, which is often associated with a traumatic event, surgical intervention will likely be required. Surgical intervention typically involves either internal fixation of the fractured bone or the femoral stem implant is replaced with a new implant. Users should consider potential risk factors for PFF that are available in published literature when evaluating its use during this time, such as those presented by Lamb et al.⁵ and ensure patients are informed appropriately on potential risks associated with the CPT Hip System Femoral Stem 12/14 Neck Taper. The potential modifiable risk factors include:

- Increasing age
- Intra-operative fracture
- Increasing stem offset
- Increasing head size
- Low viscosity bone cement

Additional Information on Affected Product

The CPT Hip System Femoral Stem 12/14 Neck Taper has been on the market for over 20 years. During this time, it achieved and has maintained the highest possible Orthopaedic Data Evaluation Panel (ODEP) rating of 15A*¹, as well as provided 96.4% all-cause survivorship at 10 years in the latest United Kingdom National Joint Registry (UK NJR) annual report.²

Product Phase Out

Zimmer Biomet has also made the decision to phase out the CPT Hip System Femoral Stem 12/14 Neck Taper, with a target of December 2024 to stop sales in your market. The product has successful long-term survivorship and may continue to be implanted during this phase out period.

Zimmer Biomet recognizes the importance of a well-planned phase out and transition when changing hip stem brands and has obtained feedback on preferred phase out periods and approaches from several surgeons who use the CPT Hip System Femoral Stem 12/14 Neck Taper. We will be taking actions to aid surgeons in transitioning to new hip stem brands, such as offering dedicated training opportunities in your market on alternative Zimmer Biomet brands and adjusting supply plans accordingly.

Hospital Responsibilities:

1. Review this Field Safety Notice and ensure that affected personnel, including surgeons, are aware of the contents.
 - a. There are no specific patient monitoring instructions related to this Field Safety Corrective Action that are recommended beyond your existing follow-up schedule.
2. If you currently use CPT Hip System Femoral Stem 12/14 Neck Taper, please coordinate with your local Zimmer Biomet representative on a transition plan for this product.
3. Complete **Attachment 1 – Certificate of Acknowledgement** and send it to fieldaction.uk@zimmerbiomet.com. This form must be returned even if you no longer use CPT Hip System Femoral Stem 12/14 Neck Taper.
4. Retain a copy of **Attachment 1 – Certificate of Acknowledgement** with your Field Safety Corrective Action records in the event of a compliance audit of your facility's documentation.
5. If you have further questions or concerns after reviewing this Field Safety Notice, please contact your local Zimmer Biomet representative.

Surgeon Responsibilities:

1. Review this Field Safety Notice for awareness of the contents.
 - a. There are no specific patient monitoring instructions related to this Field Safety Corrective Action that are recommended beyond your existing follow-up schedule.
2. Consider the benefits and risks of implanting the CPT Hip System Femoral Stem 12/14 Neck Taper on an individual patient basis and ensure patients are informed appropriately on the increased risk of PFF.
3. To assist with communicating the increased risk of PFF to prospective patients during the phase out period, **Attachment 3 – Prospective Patient Information** may be provided to patients considering CPT Hip System Femoral Stem 12/14 Neck Taper.
4. Zimmer Biomet is in the process of updating the IFU to reflect the risk of PFF. The latest IFU can be viewed electronically at labeling.zimmerbiomet.com by entering one of the material numbers from **Attachment 2 – Affected Product List** in the search field.
5. If you have further questions or concerns after reviewing this Field Safety Notice, please contact your local Zimmer Biomet representative.


Other Information

This Field Safety Corrective Action and related communications to users were reported to all relevant Competent Authorities and Notified Bodies as required under the applicable regulations for Medical Devices per Regulation (EU) 2017/745 and guidance document MDCG 2023-3. The undersigned confirms that this Field Safety Notice has been delivered to the appropriate Regulatory Agencies. Please be aware that the names of user facilities notified are routinely provided to the Competent Authorities for audit purposes.

Please keep Zimmer Biomet informed of any adverse events associated with this product or any other Zimmer Biomet product by emailing per.uk@zimmerbiomet.com.

We would like to thank you for your co-operation in advance and regret any inconvenience caused by this Field Safety Corrective Action.

Sincerely,



Francis Moloney, VP QA/RC EMEA

References

¹Orthopaedic Data Evaluation Panel. www.odep.org.uk/product/cpt-stem-cocr/

²National Joint Registry. 20th Annual Report. 2023. www.njrcentre.org.uk

³Palan J, Smith MC, Gregg P et al. The influence of cemented femoral stem choice on the incidence of revision for periprosthetic fracture after primary total hip arthroplasty. *Bone Joint J.* 2016;98-B:1347-54. doi: 10.1302/0301-620X.98B10. PMID: 27694588.

⁴Jain S, Lamb JN, Pandit H. Cemented femoral stem design and postoperative periprosthetic fracture risk following total hip arthroplasty. *Bone Joint J.* 2024 Jan 1;106-B(1):11-15. doi: 10.1302/0301-620X.106B1.BJJ-2023-0587.R1. PMID: 38160687.

⁵Lamb JN, Jain S, King SW, West RM, Pandit HG. Risk Factors for Revision of Polished Taper-Slip Cemented Stems for Periprosthetic Femoral Fracture After Primary Total Hip Replacement: A Registry-Based Cohort Study from the National Joint Registry for England, Wales, Northern Ireland and the Isle of Man. *J Bone Joint Surg Am.* 2020 Sep 16;102(18):1600-1608. doi: 10.2106/JBJS.19.01242. PMID: 32604382.

⁶Pandit HG et al. Postoperative periprosthetic femoral fracture is the leading cause of major reoperation in the United Kingdom following primary total hip replacement: A study using national health data linked to National Joint Registry. Unpublished.



ATTACHMENT 1 - Certificate of Acknowledgement

IMMEDIATE RESPONSE REQUIRED – TIME SENSITIVE ACTION NEEDED

**Affected Product: CPT® Hip System Femoral Stem 12/14 Neck Taper
Field Safety Corrective Action Reference Number: ZFA-2024-00121**

Do you or your facility currently use the CPT Hip System Femoral Stem 12/14 Neck Taper?

Yes, I/we currently use this product. **No**, I/we no longer use this product.

Comments (as applicable):

Hospital Acknowledgement

By signing below, I acknowledge that I have received, read, and understand the contents of this Field Safety Notice. All required activities are complete or are being completed.

Printed Name: _____ **Signature:** _____

Title: _____ **Telephone:** _____ **Date:** _____

Facility Name: _____

Facility Address: _____

City: _____ **Country:** _____ **ZIP/Post Code:** _____

ATTACHMENT 2 - Affected Product List

Material Number	GTIN Number	Material Description
00-8114-000-00	00889024145733	Size 0 105 mm Stem Length Standard Offset
00-8114-000-10	00889024145740	Size 0 105 mm Stem Length Extended Offset
00-8114-001-00	00889024145757	Size 1 130 mm Stem Length Standard Offset
00-8114-001-10	00889024145764	Size 1 130 mm Stem Length Extended Offset
00-8114-002-00	00889024145771	Size 2 130 mm Stem Length Standard Offset
00-8114-002-10	00889024145788	Size 2 130 mm Stem Length Extended Offset
00-8114-002-30	00889024145801	Size 2 130 mm Stem Length Extra Extended Offset
00-8114-003-00	00889024145818	Size 3 130 mm Stem Length Standard Offset
00-8114-003-10	00889024145825	Size 3 130 mm Stem Length Extended Offset
00-8114-003-30	00889024145849	Size 3 130 mm Stem Length Extra Extended Offset
00-8114-004-00	00889024145856	Size 4 130 mm Stem Length Standard Offset
00-8114-004-10	00889024145863	Size 4 130 mm Stem Length Extended Offset
00-8114-004-30	00889024145900	Size 4 130 mm Stem Length Extra Extended Offset
00-8114-005-00	00889024145917	Size 5 130 mm Stem Length Standard Offset
00-8114-005-10	00889024145924	Size 5 130 mm Stem Length Extended Offset
00-8114-005-30	00889024145931	Size 5 130 mm Stem Length Extra Extended Offset
00-8114-040-00	00889024145962	Extra Small 85 mm Stem Length
00-8114-050-00	00889024145979	Small 95 mm Stem Length
00-8114-002-18	00889024145795	Size 2 180 mm Stem Length Standard Offset
00-8114-003-18	00889024145832	Size 3 180 mm Stem Length Extended Offset
00-8114-004-20	00889024145870	Size 4 200 mm Stem Length Extended Offset
00-8114-004-23	00889024145887	Size 4 230 mm Stem Length Extended Offset
00-8114-004-26	00889024145894	Size 4 260 mm Stem Length Extended Offset
00-8114-012-18	00889024145948	Size 2 180 mm Stem Length Valgus Neck
00-8114-013-18	00889024145955	Size 3 180 mm Stem Length Valgus Neck



ATTACHMENT 3 – Prospective Patient Information

Your surgeon may be considering the CPT Femoral Stem for your hip replacement surgery. This additional information is being provided to communicate a risk that is not identified in the current instructions for use for the CPT Femoral Stem.

All hip replacements have potential risks associated with their use, including fracture of the femoral bone after surgery. This is typically caused by a traumatic event, such as a fall. The CPT Femoral Stem does not list this risk of bone fracture around the implant in the current instructions for use.

Based on a review of patients in the United Kingdom implanted with the CPT Femoral Stem, approximately 1.4% of patients experienced a femoral bone fracture after surgery. This rate of occurrence is higher than that of similar femoral stem implants.

If fracture of the femoral bone occurs, an additional surgery is typically required to repair the bone and potentially replace the femoral stem implant.

The CPT Femoral Stem has been on the market for over 20 years and has strong long-term survivorship. More than 96% of patients receiving a CPT Femoral Stem still have the implant after 10 years.

As with any surgery, you are encouraged to discuss the risk benefit of using the CPT Femoral Stem with your surgeon.